



INSTYTUT TECHNOLOGII ELEKTRONOWEJ

Al. Lotników 32/46, 02-668 Warsaw, Poland

<http://www.ite.waw.pl>

DKZ/13/2008

Warsaw, 04.07.2008 r.

Re: The open procedure for delivery of the spectroscopic ellipsometer.

With regard to the following questions:

1. *Do you agree for application of alternative device instead of monochromator in order to fulfill your conditions of the Terms of Reference concerning the constant measurement resolution (0.1 nm)?*
2. *Do you agree for application of a manual goniometer with positioning angle accuracy of 0.02°? The required value in terms of reference (0,01°) is not achievable by manual goniometers.*
3. *In connection with our questions, We would like to ask you for extension of your tender deadline for one week, till 16.07.2008.*

please be inform that we do not allow for any changes.

In connection with the above statements of the Terms of Reference remains without changes.

Zbigniew Poznański

Deputy Director ITE

Director tel. +48 22 548 77 00; +48 22 843 30 93
Deputy Director tel. +48 22 548 78 00
Vice-Director for Research and Development tel. +48 22 548 77 60
Vice-Director for Division of Silicon MNT, tel. +48 22 716 59 90
Vice-Director for Division in Krakow tel. +48 12 656 51 83
Chief Accounting Officer tel. +48 22 548 79 00

ISO 9001



Fax: +48 22 847 06 31
NIP: 525-000-91-11

Bank Account:
BRE Bank S.A., II Oddział Korporacyjny w Warszawie
47 1140 1977 0000 5580 4500 1001